



ILLINOIS TEST LAB

---

ANSI N42.17A, Test Report  
November 6, 2009

Instrument: Model 9-3 and Model 9-4

Serial Number: Model 9-3: 246808, 255485 and Model 9-4: 265867, 265868

Test Dates (inclusive): 12/08/04 – 12/29/04

Test Performed: ANSI N42.17A-2003:

6.4 Geotropism

7.6 Photon radiation overload

Results:

6.4 Geotropism-variations of less than 2% were noted for two instruments per the test method specified

7.6 Photon radiation overload-no saturation effects were noted at radiation fields up to 2000 R/hr for one minute on any of the five ranges

Conclusion:

All instruments passed the requirements of ANSI N42.17A-2003 for these tests.

Rollie Cantu, Engineering Technician